

<b>SITE LOCATION</b>		<b>BILL TO:</b>	
City of Scottsdale, Police Department 7601 East McKellips Road, Scottsdale, AZ 85257 Mr. Allan Kosecki, pkosecki@scottsdaleaz.gov		Full Spectrum	
Service Rep: Joe Galvez	Service Type: PM	Quality Level: Standard	
Billing Type: Non-Billable	Payment Method: Service Contract	PO#: N/A	
System ID	Manufacturer	Model	Serial#
GC w/ HS			US14173023
Additional Instruments	7890 GC S/N CN14160045		
<b>PROBLEM DESCRIPTION:</b>			
Agilent Headspace GC/FID Annual Preventative Maintenance			
<b>WORK PERFORMED:</b>			
Lowered the HS and GC/FID temperatures. Turned off flow of the Hydrogen carrier gas. Cleaned all dust off and around GC and headspace instruments. Changed the headspace sampling needle and sample loop. Returning all temps and flows to original settings. Injected a test sample. There was no response. Located small leak at the 6-port valve. Performed another test injection with acceptable results. System is operating properly.			

FSA Part# (Inventory)	Description (manual)	Qty	Price Each	Total
	7697A Headspace Sample Loop (1 mL)	1	\$444.00	\$444.00
	7697A Headspace Sample Needle	1	\$470.00	\$470.00

Parts Total: \$914.00

HOURS	RATE	TOTAL	
Labor Hours: 3.0	Labor Rate: \$300	Labor Total:	\$900
Travel Hours: 2.0	Travel Rate: \$300	Travel Total:	\$600
	Extended Travel: \$0	Misc. S&H:	\$0.00
	<b>NOT AN INVOICE, TAXES NOT INCLUDED</b>	Total Labor:	\$1500
		Report Total:	\$2414.00
		Credit Amount:	\$2414.00
		<b>Balance Due:</b>	<b>\$0.00</b>