



ARIZONA DEPARTMENT OF PUBLIC SAFETY  

---

INTOXILYZER MAINTENANCE REPORT

CR NO. I800018790007

Page 1 of 1

AGENCY           Scottsdale Police Department  
                      Scottsdale, AZ 85251  
OFFICER          JUDD, #1232  
DATE             November 06, 2019



**INSTRUMENT**

80001879

**INSTRUMENT LOCATION**

Scottsdale Police Department

**MAINTENANCE REQUESTED**

Repair and Preventative Maintenance

**MAINTENANCE PERFORMED**

***Parts Replaced:***

Breath Hose, Source

***Preventative Maintenance Performed:***

Inspection, Calibration, Verification, Adjustment

A handwritten signature in black ink, appearing to read "Joseph M Slowinski".

---

JOSEPH M SLOWINSKI, #6101, Electronics Technician  
Central Regional Crime Laboratory  
2102 W. Encanto Blvd., Phoenix, AZ 85009  
602-223-2159   jslowinski@azdps.gov



ARIZONA DEPARTMENT OF PUBLIC SAFETY  

---

INTOXILYZER MAINTENANCE REPORT

CR NO. I800014640007

Page 1 of 1

AGENCY           Scottsdale Police Department  
                      Scottsdale, AZ 85251  
OFFICER          JUDD, #1232  
DATE             November 05, 2019



**INSTRUMENT**

80001464

**INSTRUMENT LOCATION**

Scottsdale Police Department

**MAINTENANCE REQUESTED**

Repair and Preventative Maintenance

**MAINTENANCE PERFORMED**

*Parts Replaced:*

Breath Hose, End Block, Source

*Preventative Maintenance Performed:*

Inspection, Calibration, Verification, Adjustment

*David K. Worthen*

---

DAVID K. WORTHEN, #4871, Forensic Scientist  
Central Regional Crime Laboratory  
2102 W. Encanto Blvd., Phoenix, AZ 85009  
602-223-2415   dworthen@azdps.gov